



ALBAcheck® - BGS Anti-D 0.05 IU/mL

Human Polyclonal / Indirect Agglutinin

For Sensitivity Control of the
Anti-Human Globulin Test

REF Z263



Storage temperature limitation (2°C– 8°C)



In vitro diagnostic medical device



Consult instructions for use



Harmful



Manufacturing



Product Code

INTENDED PURPOSE

ALBAcheck®-BGS Anti-D (0.05 IU/mL) is intended for use as a sensitivity control of the anti-human globulin test.

REAGENT DESCRIPTION

This reagent has been prepared from plasma collected from blood donors. Each individual donation contains IgG anti-D of less than 10 IU/mL. Some donations may contain IgG antibodies other than anti-D as minor contaminants.

Conversion to serum was achieved by the addition of calcium chloride and, when necessary thrombin. Excess calcium was removed by the addition of sodium oxalate. The formulation also contains 0.1% (w/v) sodium azide.

The volume delivered by the reagent dropper bottle is approximately 40 µL; bearing this in mind, care should be taken to ensure that appropriate serum: cell ratios are maintained in all test systems.

This reagent complies with the requirements of Directive 98/79/EC on *in vitro* Diagnostic Medical Devices and the recommendations contained in the Guidelines for Blood Transfusion Services in the United Kingdom.

STORAGE CONDITIONS

The reagent should be stored at 2-8 °C. Do not use if turbid. Do not dilute. The reagent is stable until the expiry date stated on the product label.

PRECAUTIONS FOR USE AND DISPOSAL

This reagent contains 0.1% (w/v) sodium azide (EC No.247-852-1) and is classified as harmful. R22 Harmful if swallowed.

Sodium azide may react with lead and copper plumbing to form explosive compounds. If discarded into sink, flush with a large volume of water to prevent azide buildup.

CAUTION: SOURCE MATERIAL FROM WHICH THIS PRODUCT IS DERIVED WAS FOUND NON-REACTIVE FOR HBsAg, ANTI-HIV 1/2 AND ANTI-HCV. NO KNOWN TEST METHODS CAN OFFER ASSURANCE THAT PRODUCTS DERIVED FROM HUMAN BLOOD WILL NOT TRANSMIT INFECTIOUS DISEASE. APPROPRIATE CARE SHOULD BE TAKEN IN THE USE AND DISPOSAL OF THIS PRODUCT.

This reagent is for *in vitro* professional use only.

TEST PROCEDURES

This reagent has been standardised for use by the techniques described below and therefore its suitability for use in other techniques cannot be guaranteed.

RECOMMENDED TECHNIQUES

DiaMed - ID Micro Typing System

Additional Materials and Reagents Required

- ID card "LISS/Coombs"
- ID Diluent 2 - modified LISS
- ID centrifuge
- ID Dispenser
- ID Pipettor
- ID Pipettor tips
- Tubes for cell suspensions
- ID working table
- ID incubator

Test Procedure

- Prepare cell suspension - 12.5 µL packed cells + 1.0 mL ID Diluent 2.
- Add 50 µL cells to the appropriate microtubes.
- Add 25 µL of the Anti-D control to each microtube.
- Incubate for 15 minutes at 37 °C in the ID incubator.
- Centrifuge in the ID centrifuge for 10 minutes.
- Read reactions.

Ortho BioVue® System

Additional Materials and Reagents Required

- Anti-IgG,-C3 cards
- Ortho BLISS
- Ortho BioVue Centrifuge
- Ortho BioVue Incubator
- Ortho BioVue Pipette
- Ortho BioVue Pipette tips
- Ortho BioVue workstation
- Tubes for cell suspension

Test Procedure

- Prepare cell suspension – 10 µL packed cells + 0.8 mL BLISS.
- Add 50 µL of test cells to appropriate microtubes.
- Add 40 µL anti-D control to each microtube.



INTRODUCTION

ALBAcheck®-BGS is a registered trademark for a group of products manufactured by Alba Bioscience for use as controls in blood group serology tests.

Described in 1945 by Coombs, Mourant and Race, the Coombs, or anti-human globulin technique, remains one of the most important and widely used techniques in blood group serology. A multiplicity of variables can affect the outcome of antiglobulin tests and consequently it is good laboratory practice to employ a procedure which can be used to indicate the sensitivity and limitations of a test system.

This Anti-D (0.05 IU/mL) can be used to evaluate the sensitivity of the antiglobulin test procedure performed in the laboratory.

INTERPRETATION OF LABELLING SYMBOLS



Batch code



Use by (YYYY-MM-DD)

- . Incubate for 15 minutes at 37 °C in Ortho BioVue incubator.
- . Centrifuge in Ortho BioVue centrifuge for 5 minutes.
- . Read reactions.

INTERPRETATION OF RESULTS

Agglutination = positive test result
No agglutination = negative test result

QUALITY CONTROL

This is a quality control reagent and its satisfactory performance when used by the recommended techniques represents an adequate level of control.

PERFORMANCE LIMITATIONS

R₁r samples will exhibit variable degrees of RhD antigen expression.

Improper techniques may invalidate the results obtained with this product.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials and omission of test reagents.

SPECIFIC PERFORMANCE CHARACTERISTICS

THIS REAGENT SHOULD ONLY BE USED IN THE APPLICATIONS DESCRIBED AND IS NOT SUITABLE FOR RhD GROUPING.

DATE OF ISSUE

2016-03-04

For further information or advice please contact your local distributor.



Alba Bioscience
Ellen's Glen Road
Edinburgh
Scotland, UK
EH17 7QT

Tel No: +44 (0) 131 658 5700
Fax No: +44 (0) 131 672 3026
E-Mail: customer.serviceEU@quotientbd.com