

BLOOD GROUPING REAGENT

Anti-C ALBAclone®

(Human/Murine Monoclonal)
For Tube Technique

REF Z064U

- Meets FDA potency requirements
- Discard if turbid
- Preservative: 0.1% (w/v) sodium azide

CAUTIONS: THE ABSENCE OF ALL VIRUSES HAS NOT BEEN DETERMINED. THIS PRODUCT HAS COMPONENTS (DROPPER BULBS) CONTAINING DRY NATURAL RUBBER.

INTERPRETATION OF LABELING SYMBOLS

LOT

Batch code



Use by (YYYY-MM-DD)

REF

Product Code



Storage temperature limitation (2–8 °C)

IVD

In vitro diagnostic medical device



Consult instructions for use

www.quotientbd.com



Manufacturer

INTENDED USE

This Anti-C reagent is for the *in vitro* detection and identification of the human C blood group antigen by direct agglutination.

SUMMARY AND EXPLANATION

Since the description of the RhD antigen by Levine and Stetson in 1939, more than 40 other Rh antigen complexes have been identified. With the exception of C, c, E and e, and perhaps C^w, few of these antigens or their corresponding antibodies are encountered in routine testing. Rh antigens are controlled by a series of closely linked loci on chromosome 1, the genetic contribution from each parent being inherited as a haplotype e.g. Cde, cDE etc. Used separately, anti-Rh blood grouping reagents will indicate whether an individual expresses the corresponding antigen - an essential procedure in the selection of blood for transfusion of patients with Rh antibodies.

Testing red blood cell samples with anti-C, anti-D, anti-E, anti-c and anti-e will disclose the Rh phenotype from which the most probable genotype may be deduced. Knowing the probable paternal genotype can be of value in the management of RhD hemolytic disease of the fetus and newborn where R_{2r} infants are likely to be more severely affected than R_{1r} infants. Probable genotype information can also be useful in establishing antibody specificity and in selecting blood for transfusion of patients with Rh antibodies.

PRINCIPLE OF THE TEST

When used by the recommended technique, this reagent will cause agglutination (clumping) of red blood cells carrying the C antigen. Lack of agglutination demonstrates the absence of the C antigen.

REAGENT DESCRIPTION

The main component of this reagent is IgM antibody derived from the *in vitro* culture of the IgM secreting human/mouse heterohybridoma of cell line MS-24:

Product Name	Product Code	Cell Line
Anti-C	Z064U	MS-24

The formulation contains bovine material, potentiators and 0.1% (w/v) sodium azide.

The volume delivered by the reagent bottle dropper is approximately 40 µL. Care should be taken to ensure that appropriate serum to cell ratios are maintained in all test systems.

WARNINGS AND PRECAUTIONS

For *in vitro* diagnostic use only
Products should be used by qualified personnel
Do not use beyond the expiration date.

Do not use if turbid

Do not dilute

The format of the expiration date is expressed as YYYY-MM-DD (Year-Month-Day)

This reagent contains 0.1% (w/v) sodium azide. Sodium azide may be toxic if ingested and may react with lead and copper plumbing to form explosive compounds. If discarded into a sink, flush with a large volume of water to prevent azide build-up.

This reagent is of animal origin (murine and bovine), therefore care must be taken during use and disposal as there is a potential infection risk.

CAUTION: ALL BLOOD PRODUCTS SHOULD BE TREATED AS POTENTIALLY INFECTIOUS. SOURCE MATERIAL FROM WHICH THIS PRODUCT WAS DERIVED WAS FOUND NEGATIVE WHEN TESTED IN ACCORDANCE WITH CURRENT FDA REQUIRED TESTS. NO KNOWN TEST METHODS CAN OFFER ASSURANCE THAT PRODUCTS DERIVED FROM HUMAN BLOOD WILL NOT TRANSMIT INFECTIOUS AGENTS.

The bovine material used in the manufacture of this reagent was collected in a USDA approved facility.

Monoclonal antibodies exhibit a high degree of potency, avidity and specificity. When using such antibodies, great care should be taken to avoid cross contamination.

This product has components (dropper bulbs) containing dry natural rubber.

STORAGE CONDITIONS

The reagent should be stored at 2–8 °C.

SPECIMEN COLLECTION AND PREPARATION

Specimens should be collected by a standard collection technique. The specimen should be tested as soon as possible after collection. If testing is delayed, the specimen should be stored at refrigerated temperatures.

Clotted samples, or those collected in EDTA, should be tested within fourteen days from collection. Donor blood collected in ACD, CPD, CPDA -1, CP2D, CP2D with AS-3, and CPD with AS-1 and CPD with AS-5 may be tested until the expiration date of the donation.

Special care should be taken if hemolyzed samples must be tested. Grossly icteric or contaminated blood specimens should not be used.

MATERIALS

Material provided

- ALBAclone® Anti-C

Materials required but not provided

- Isotonic saline
- Reagent red blood cells suitable for the control of Anti-C
- 10 x 75 mm or 12 x 75 mm glass test tubes
- Pipettes
- Optical aid (optional)
- Centrifuge
- Timer
- Heating block/waterbath

PROCEDURE

General Information

NOTE: This reagent has been standardized for use by the technique described below and therefore its suitability for use by other techniques cannot be guaranteed.

When a test is required to be incubated for a specific time period, a timer should be used.

When using supplemental testing equipment (i.e. centrifuge), follow the procedures that are contained in the operator's manual provided by the device manufacturer.

RECOMMENDED TECHNIQUE

Tube Technique – 5-15 Minute Incubation/Spin

1. Prepare a 2-4% suspension of red blood cells in isotonic saline solution (Reagent Red Blood Cells may be used directly from the vial or according to the manufacturer's instructions.)
2. Add 1 drop of blood grouping reagent to a glass test tube.
3. Add 1 drop of red blood cell suspension. Steps 2 and 3 may be performed in either order.
4. Mix the contents of the test tube and incubate at 37 °C ± 1 °C for 5-15 minutes.
5. Centrifuge the test tube.

NOTE: Suggested centrifugation: 900-1000 g (approx. 3400 rpm) for 10 seconds or a time and speed appropriate for the centrifuge used that produces the strongest reaction of antibody with antigen-positive red blood cells, yet allows easy re-suspension of antigen-negative red blood cells.

6. After centrifugation, gently shake the tube to dislodge the cell button from the bottom and immediately observe macroscopically for agglutination. Negative reactions may be examined with an optical aid.
7. Record results.

STABILITY OF REACTION

Test results should be read and interpreted immediately after centrifugation. Delays may cause dissociation of antigen-antibody complexes resulting in weak positive or false negative reactions.

INTERPRETATION OF RESULTS

Agglutination = positive test result
No agglutination = negative test result

QUALITY CONTROL

Quality control of reagents is essential and should be performed on the day of use and in accordance with local, state and federal regulations.

C+c+ red blood cells should be used as a positive control.
C-c- red blood cells should be used as a negative control.

PERFORMANCE LIMITATIONS

Heating blocks and waterbaths promote better heat transfer and are recommended for 37 °C tests, particularly where the incubation period is 30 minutes or less.

The expression of certain red blood cell antigens may diminish in strength during storage, particularly in EDTA and clotted samples. Better results will be obtained with fresh samples.

Gently re-suspend tube tests before reading. Excessive agitation may disrupt weak agglutination and produce false negative results.

Excessive centrifugation can lead to difficulty in re-suspending the cell button, while inadequate centrifugation may result in agglutinates that are easily dispersed.

False positive or false negative results can occur due to contamination of test materials, improper reaction temperature, improper storage of materials, omission of test reagents and certain disease states.

Suppressed or weak expression of blood group antigens may give rise to false negative reactions.

SPECIFIC PERFORMANCE CHARACTERISTICS

Prior to release, each lot of ALBAclone® Anti-C is tested by FDA recommended methods against a panel of antigen-positive and antigen-negative red blood cells to ensure suitable reactivity.

Comparator Study Results

During comparator studies (data on file at Alba Bioscience Limited), blood samples were tested with ALBAclone® Anti-C (Monoclonal) as follows:

Anti-C		Comparator Reagent			One-sided 95% Exact lower confidence limit
		Positive	Negative	Total	
Trial Reagent	Positive	703	0	703	
	Negative	0	336	336	
	Total	703	336	1039	
Positive Percent Agreement				100	0.99
Negative Percent Agreement				100	0.99

In performance evaluation studies, 1039 samples were tested with ALBAclone® Anti-C (Monoclonal). The positive percent agreement at the one-sided 95% exact lower confidence limit was 0.99 for agglutination tests based on a comparison of interpreted results. The negative agreement at the one-sided 95% exact lower confidence limit was 0.99 for agglutination tests based on a comparison of interpreted results.

Results were evaluated against comparable FDA approved products using the appropriate methods for the comparators.

Precision Study Results

As part of the performance evaluation, precision and lot to lot studies were performed using multiple operators, days and runs to confirm repeatability and reproducibility of test results

in the same run, day and with the same operator and between runs, days and operators. The study took into account variables such as days of the week, times of day and supplementary reagents used in testing. There were no discordant results; all antigen positive test outcomes generated unequivocal positive reactions and antigen negative test outcomes generated unequivocal negative reactions.

BIBLIOGRAPHY

1. Roback JD, Grossman BJ, Harris T, *et al*: AABB Technical Manual, 18th ed. AABB, 2014
2. AABB Standards Program Committee: Standards for Blood Banks and Transfusion Services, 29th ed. AABB, 2014
3. Reid ME, Lomas-Francis C, Olsson ML: The Blood Group Antigen FactsBook, ed 3. Academic Press, 2012

DATE OF ISSUE

2019-02-28

US Distributor

Quotient
301 S. State Street
S-204
Newtown
PA 18940
USA

Customer Service Tel: 1-888-284-1901
Product Technical Support Tel: 1-888-228-1990
Customer Service Fax: 1-888-694-5208
E-Mail: customer_serviceUS@quotientbd.com
Web: <http://us.quotientbd.com>



Alba Bioscience Limited
James Hamilton Way
Penicuik
EH26 0BF
UK

US License 1807

© Alba Bioscience Limited 2019

Z064UPI/02



QUOTIENT